

## COMPARISON TABLE

COMPARED PRODUCTS	M-BASIC	M-DIAGNOSTIC	M-SCIENCE
			
<b>POWER SUPPLY</b>	230V 50/60Hz; 120V 50/60Hz (option 100, 110, 127V 50/60Hz)	230V-50/60Hz; 120V-50/60Hz (option 100, 110, 127V 50/60Hz)	230V-50/60Hz; 120V-50/60Hz (option 100, 110, 127V 50/60Hz)
<b>POWER [W]</b>	120	190	230
<b>CAPACITY [ML]</b>	8 x 15; 2 x 50; 2 x 60 PRP	500, 4 x CYTO	100, (24 capillaries 75 mm)
<b>MAX. RPM [M<sup>-1</sup>]</b>	300 ÷ 4000, step 100 RPM	90 ÷ 6 000, step 1 RPM	90 ÷ 18 000, step 1 RPM
<b>MAX. RCF [X G]</b>	2 469	4 830	24 270
<b>SETTING TIME</b>	1 ÷ 99 min, step 1 min, ∞ (HOLD)	1s ÷ 99h 59min 59s, ∞, step 1s	1s ÷ 99h 59min 59s, ∞, step 1s
<b>DIMENSIONS H X W X D [MM]</b>	300 x 365 x 450	299 x 357 x 451	299 x 357 x 451
<b>HEIGHT WITH OPEN LID [MM]</b>	572	572	572

### "MPW MED. INSTRUMENTS" SPÓŁDZIELNIA PRACY

ul. Boremlowska 46, 04-347 Warszawa  
tel.: +48 22 610 56 21 / fax: +48 22 610 55 36  
mail: mpw@mpw.pl / www.mpw.pl

REGON: 000401029  
KRS: 0000128941  
VAT UE: PL 5250001753  
BDO 000042924

Alior Bank S.A.  
16 2490 0005 0000 4530 5256 8362 **PLN**  
PL 44 2490 0005 0000 4600 2441 7250 **EUR**

<b>WEIGHT [KG]</b>	~ 21/23*	~ 22/23,7*	~ 20/21,5*
<b>MIN. ACCELERATION TIME [S]</b>	35	5	14
<b>MIN. BRAKING TIME [S]</b>	33	11	12
	* weight of the centrifuge depends on the voltage version	* weight of the centrifuge depends on the voltage version	* weight of the centrifuge depends on the voltage version

**"MPW MED. INSTRUMENTS" SPÓŁDZIELNIA PRACY**

ul. Boremlowska 46, 04-347 Warszawa  
 tel.: +48 22 610 56 21 / fax: +48 22 610 55 36  
 mail: mpw@mpw.pl / www.mpw.pl

REGON: 000401029  
 KRS: 0000128941  
 VAT UE: PL 5250001753  
 BDO 000042924

Alior Bank S.A.  
 16 2490 0005 0000 4530 5256 8362 **PLN**  
 PL 44 2490 0005 0000 4600 2441 7250 **EUR**