

DECLARATION OF CONFORMITY

Product name: **Laboratory centrifuge MPW-56**

Product type: **Laboratory centrifuge**

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product classification on the basis of the Directive 98/79/EC: **Non classified to list A or B and not for self-testing.**

Product complies with the requirements:

• Directive 98/79/EC (IVD), including the requirements of harmonized standards:

EN 15223-1:2016

EN ISO 18113-3:2011

EN 13612:2002

EN 61326-2-6:2006

EN 13612:2002/AC:2002

EN 61010-2-101:2002

EN 13975:2003

EN 62304:2006

EN ISO 14971:2012

EN 62304:2006/AC:2008

EN ISO 18113-1:2011

EN 62366:2008

• selected harmonized standards of Directive 2014/35/UE (LVD):

EN 61010-1:2010

EN 61010-2-020:2006

• directive 2014/30/UE (EMC).

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
w Warszawie

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
Warsaw, 46 Boremlowska Street
*applies Quality Management System in line with
PN-EN ISO 9001:2015, PN-EN ISO 13485:2016*
Certifying authority:

Członek Zarządu
Wojciech Anisiewicz

PREZES ZARZĄDU

mgr Łukasz Satański

