

DECLARATION OF CONFORMITY

Product name: **Laboratory centrifuge MPW-55**

Product type: **Laboratory centrifuge**

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product classification on the basis of the Directive 98/79/EC: Non classified to list A or B and not for self-testing.

Product complies with the requirements:

• **Directive 98/79/EC (IVD), including the requirements of harmonized standards:**

EN 15223-1:2016	EN ISO 18113-3:2011
EN 13612:2002	EN 61326-2-6:2006
EN 13612:2002/AC:2002	EN 61010-2-101:2002
EN 13975:2003	EN 62304:2006
EN ISO 14971:2012	EN 62304:2006/AC:2008
EN ISO 18113-1:2011	EN 62366:2008

• **selected harmonized standards of Directive 2014/35/UE (LVD):**

EN 61010-1:2010	EN 61010-2-020:2006
-----------------	---------------------

• **directive 2014/30/UE (EMC).**

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
w Warszawie

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
Warsaw, 46 Boremlowska Street
applies Quality Management System in line with
PN-EN ISO 9001:2015, PN-EN ISO 13485:2016
Certifying authority:

Członek Zarządu
Wojciech Anisiewicz

PREZES ZARZĄDU
mgr Łukasz Satański

