

DECLARATION OF CONFORMITY

Product **CYTO insert cytology rotor, type 610**

Type **Type 610**

Catalogue number **16610**

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product classification on the basis of the Directive 98/79/EC **Non classified to list A or B and not for self-testing**

Product complies with the requirements:

• **Directive 98/79/EC (IVD), including the requirements of harmonised standards:**

EN 15223-1:2016

EN ISO 14971:2012

EN 13612:2002

EN ISO 18113-1:2011

EN 13612:2002/AC:2002

EN ISO 18113-3:2011

EN 13975:2003

EN 62366:2008

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
w Warszawie

„MPW MED. INSTRUMENTS”
SPÓŁDZIELNIA PRACY
Warsaw, 46 Boremlowska Street
applies Quality Management System in line with
PN-EN ISO 9001:2015, PN-EN ISO 13485:2016
Certifying authority:

Członek Zarządu
Wojciech Anisiewicz

PREZES ZARZĄDU

mgr Łukasz Satański



Warsaw, 2018.09.15

no. 10.610.04.en