MPW MED. INSTRUMENTS

DECLARATION OF CONFORMITY

Product name: **Laboratory centrifuge MPW-56**

Product type: **Laboratory centrifuge**

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product classification on the basis of the Directive 98/79/EC: Non classified to list A or B and not for self-testing.

Product complies with the requirements:

· Directive 98/79/EC (IVD), including the requirements of harmonized standards:

EN 15223-1:2016
EN 13612:2002
EN 13612:2002/AC:2002
EN 13975:2003
EN ISO 14971:2012
EN ISO 18113-1:2011
EN ISO 18113-3:2011
EN 61326-2-6:2006
EN 61010-2-101:2002
EN 62304:2006
EN 62304:2006/AC:2008
EN 62366:2008

· selected harmonized standards of Directive 2014/35/UE (LVD):

EN 61010-1:2010
EN 61010-1:2010/A1:2019
EN 61010-1:2010/A1:2019/AC:2019-04
EN 61010-2-020:2006

· Directive 2014/30/UE (EMC)

· Directive 2011/65/UE (RoHS 2)

"MPW MED. INSTRUMENTS"
SPÓŁDZIELNIA PRACY
Warsaw, 46 Boremlowska Street
applies Quality Management System in line with PN-EN ISO 9001:2015, PN-EN ISO 13485:2016

.. PREZESA ZARZĄDU
Wojciech Anisiewicz

PREZES ZARZĄDU
mgr Łukasz Sałański

ISO 9001 ISO 13485

Warsaw, 2021.08.26

no. 10.056.05.en